

VELVET GRIP

Disposable NITRILE GLOVES



curaden

better health for you

CERTIFICATIONS

cura **GRIP**⁺
A SWISS BRAND www.curagrip.com

INDEX:	01	MAIN PAGE	1
	02	INDEX	2
	03	CERTIFICATES	3
	04	ISO CERTIFICATES	3
		• ISO9001:2005	3
		• ISO 13485: 2016	4
		EN 455	
		• EN 455:1-4	5
		EN 374	
		• EN 374-1-2, EN ISO 374-4-5, EN 21420	10
		• CE 0197	17
		EN 1186	18
		FDA	19
		DoC	20
	05	PACKAGING	22
	06	LAST PAGE	23



MANUFACTURING ACCREDITATIONS

ISO 9001: 2015



MANUFACTURING ACCREDITATIONS

ISO 13485:2016



Certificate

**Quality Management System
EN ISO 13485:2016**

Registration No.:
Organization:

Scope: Design and Development, Manufacture and Distribution of Patient Examination Gloves

The Certification Body of TÜV Rheinland LGA Products GmbH certifies that the organization has established and applies a quality management system for medical devices. Proof has been furnished that the requirements specified in the abovementioned standard are fulfilled. The quality management system is subject to yearly surveillance.

Report No.:
Effective date: 2021-04-15
Expiry date: 2024-04-14
Issue date: 2021-04-13


TÜV Rheinland LGA Products GmbH
Tillystraße 2 • 90431 Nürnberg • Germany







Certificate


**Quality Management System
EN ISO 13485:2016**


Registration No.:
Organization:


The scope of certification also covers the following:

No.	Facility	Scope
/01		Distribution of Patient Examination Gloves
/02		Design and Development, Manufacture of Patient Examination Gloves

Report No.:
Effective date: 2021-04-15
Expiry date: 2024-04-14
Issue date: 2021-04-13


TÜV Rheinland LGA Products GmbH
Tillystraße 2 • 90431 Nürnberg • Germany







Certificate

**Quality Management System
EN ISO 13485:2016**

Registration No.:
Organization:

The scope of certification also covers the following:

/03	Design and Development, Manufacture of Patient Examination Gloves
/04	Design and Development, Manufacture of Patient Examination Gloves

Report No.:
Effective date: 2021-04-15
Expiry date: 2024-04-14
Issue date: 2021-04-13


TÜV Rheinland LGA Products GmbH
Tillystraße 2 • 90431 Nürnberg • Germany






LGAM
Precisely Right.

Business Stream Products
Certification Department

TÜV Rheinland LGA Products GmbH • 51105 Köln

Contact:
Tel: +49 811 855-0225
Mail: service@lga.com
Date: April 13, 2021

Application for:
Certificate No.:
Requirement: EN ISO 13485:2016

Dear Madam or Sir,
Enclosed please find the new certificate No. [redacted] replacing the previous certificate.
With effective date of the new certificate, the previous certificate becomes invalid.

Best regards,

Jing Zhang
Certification body

ISO 13485:2016 © TÜV Rheinland LGA Products GmbH. Distribution and application rules are part of the approval.



TÜV Rheinland
LGA Products GmbH
Am Glacis Stein
51105 Köln
Germany
Headquarter
Tillystraße 2
90431 Nürnberg
Phone: +49 811 855-0225
Fax: +49 811 855-0225
service@lga.com
www.lga.com/BSSEU
Board of Management
Dag. Ing. Jing Miller, Egon Schmitt
Det.-Stv. Dr. Jörn Schmitt
Nominierungsjahr
VAT No.: DE 311884605
Chairman of the Supervisory Board
Det.-Ing. Kai Dittmann



INTERNATIONAL STANDARDS

EN 455: 1-4

Test Report No.
dated 20 Oct 2020



Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSE Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

TESTED FOR:

TEST DATE:
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020


DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		S	400	

Lot size as specified by client: 150,000 to 500,000 pieces


METHOD OF TEST:

- EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use
Part 3: Requirements and testing for biological evaluation
 - Clause 4.4 Powder-free gloves
 - Clause 4.6 Labelling



APPROVED BY
Page 1 of 5
curaGrip

Test Report No.
dated 20 Oct 2020



RESULTS:

Sample: Nitrile Disposable Exam Gloves, Lot No. , Blue, Size S

Table 1: Results for EN 455-1:2020


Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	0	Passed

Table 2: Results for EN 455-2:2015 Clauses 4-5

Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	≥ 240	13	246	Passed
	b) Width (mm)	For Size S: 80 ± 10	13	83	Passed
5	a) Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	7.6	Passed
	b) Force at break after challenge testing (N) 7 days at 70±2°C	For nitrile examination gloves: ≥ 6.0	13	7.4	Passed


Table 3: Results for EN 455-2:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed



APPROVED BY
Page 2 of 5
curaGrip

Test Report No.
dated 20 Oct 2020



RESULTS (cont'd):


Sample: Nitrile Disposable Exam Gloves, Lot No. , Blue, Size S

Table 4: Results for EN 455-3:2015 Clause 4.4

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove	0.52 mg per glove	Passed
5.2				

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex. The labelling shall include the following or equivalent warning statement together with the symbol: (Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses.	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free.	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: "CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions".	NA
		d) for any medical glove containing natural rubber latex the product labelling shall not include: <ul style="list-style-type: none"> - any terms suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unqualified indication of the presence of allergens. 	NA
e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA		
		Inferred results	Passed



APPROVED BY
Page 3 of 5
curaGrip

Test Report No.
dated 20 Oct 2020



REMARKS:

- Labelling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

APPENDIX:

Yeo Poh Kwang
Associate Engineer

Loi Chai Yi
Engineer
Medical Health Services (NAM)

Photo 1: Nitrile Disposable Exam Gloves, Lot No.



Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.




APPROVED BY
Page 4 of 5
curaGrip

INTERNATIONAL STANDARDS

EN 455: 1-4

Test Report No.
dated 20 Oct 2020




Please note that this Report is issued under the following terms :

- This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SUD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SUD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
- The samples mentioned in this report were submitted/supplied/manufactured by the Client. TÜV SUD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacturer, composition or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SUD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on the sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SUD PSB or to the report or results furnished by TÜV SUD PSB in any advertisements or sales promotion.
- Unless otherwise stated, the tests were carried out in TÜV SUD PSB Pte Ltd, No. 1 Science Park Drive Singapore 118271.
- The tests carried out by TÜV SUD PSB and this report are subject to TÜV SUD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SUD Group.

Effective 01 September 2020




Test Report No.
dated 20 Oct 2020



Subject:

TESTED FOR:

TEST DATE:
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020


DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		M	400	

Lot size as specified by client: 150,001 to 500,000 pieces


METHOD OF TEST:

- EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use
Part 3: Requirements and testing for biological evaluation
- Clause 4.4 Powder-free gloves
- Clause 4.6 Labelling




Phone: +65 6861 1330
Fax: +65 6741 8852
E-mail: enquiry@tvsud.com
http://www.tvsud.com.sg
Co. Reg: 19920697R

Regional Head Office:
TUV SUD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118271
TUV®



Test Report No.
dated 20 Oct 2020



RESULTS:

Sample: Nitrile Disposable Exam Gloves, Lot No.

Table 1: Results for EN 455-1:2020


Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	3	Passed

Table 2: Results for EN 455-2:2015 Clauses 4-5


Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	≥ 240	13	245	Passed
	b) Width (mm)	For Size M: 95 ± 10	13	93	Passed
5	a) Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	6.9	Passed
	b) Force at break after challenge testing (N) 7 days at 170±2°C	For nitrile examination gloves: ≥ 5.0	13	7.1	Passed

Table 3: Results for EN 455-3:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 10223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed



Test Report No.
dated 20 Oct 2020



RESULTS (cont'd):


Sample: Nitrile Disposable Exam Gloves, Lot No.

Table 4: Results for EN 455-3:2015 Clause 4.4

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	1.08 mg per glove	Passed

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex;	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: "(Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses."	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free.	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: "CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions;"	NA
d) for any medical glove containing natural rubber latex the product labelling shall not include: <ul style="list-style-type: none"> - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unqualified indication of the presence of allergens; 	NA		
e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA		
		Inferred results	Passed



INTERNATIONAL STANDARDS

EN 455:1-4

Test Report No.
dated 20 Oct 2020



PSB Singapore

REMARKS:

- Labelling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

Yeo Poh Kwang
Associate Engineer

Loi Chi Yi
Engineer
Medical Health Services (NAM)

APPENDIX:




Photo 1: Nitrile Disposable Exam Gloves, Lot No.



Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.



Test Report No.
dated 20 Oct 2020





PSB Singapore


Please note that this Report is issued under the following terms:

- This report applies to the sample of the specific production equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such production equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the production equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific production equipment.
- The samples mentioned in this report were submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
- Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.
- The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 01 September 2020

Test Report No.
dated 20 Oct 2020



PSB Singapore

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

TESTED FOR:

TEST DATE:
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020



DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		L	399	

Lot size as specified by client: 150,001 to 500,000 pieces

METHOD OF TEST:

- EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use
Part 3: Requirements and testing for biological evaluation
- Clause 4.4 Powder-free gloves
- Clause 4.6 Labelling





TUV SÜD PSB
Laboratory:
TUV SÜD PSB Pte. Ltd.
No. 1 Science Park Drive
Singapore 118221

Phone: +65 6886 1333
Fax: +65 6716 8939
E-mail: enquiry@tuv-sud.com
http://www.tuv-sud.com/en/49
Co Reg: 199208639

Regional Head Office:
TUV SÜD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-15
Singapore 118221
TUV

Test Report No.
dated 20 Oct 2020



PSB Singapore

RESULTS:

Sample: Nitrile Disposable Exam Gloves, Lot No.

Table 1: Results for EN 455-1:2020



Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	3	Passed

Table 2: Results for EN 455-2:2015 Clauses 4.5

Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	≥ 240	13	243	Passed
	b) Width (mm)	For Size L: 110 ± 10	13	104	Passed
5	a) Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	7.5	Passed
	b) Force at break after challenge testing (N) 7 days at 170±2°C	For nitrile examination gloves: ≥ 6.0	13	8.1	Passed

Table 3: Results for EN 455-2:2015 Clause 7


Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed

INTERNATIONAL STANDARDS

EN 455:1-4

Test Report No. dated 20 Oct 2020




RESULTS (cont'd):
Sample: Nitrile Disposable Exam Gloves, Lot No.

Table 4: Results for EN 455-3:2015 Clause 4.4

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	Powder-free gloves	For powder-free gloves, The total quantity of powder residues shall not exceed 2 mg per glove.	0.96 mg per glove	Passed

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	NA
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex.	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: (Product) contains natural rubber latex, which may cause allergic reactions, including anaphylactic responses.	Comply
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free.	NA
		c) sterile powdered gloves shall be labelled with the following or equivalent: 'CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions.'	NA
d) for any medical glove containing natural rubber latex the product labelling shall not include: <ul style="list-style-type: none"> - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unqualified indication of the presence of allergens; 	NA		
e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA		
Inferred results			Passed



Test Report No. dated 20 Oct 2020



REMARKS:

- Labelling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

Yeo Poh Kwang Associate Engineer

Eng Din Yi Engineer Medical Health Services (NAM)

APPENDIX:



Photo 1: Nitrile Disposable Exam Gloves, Lot No.





Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.





Test Report No. dated 20 Oct 2020




Please note that this Report is issued under the following terms:

- This report applies to the sample of the specific production equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SUD PSB approves, recommends or endorses the manufacturer, supplier or user of such production equipment, or that TÜV SUD PSB in any way 'guarantees' the later performance of the production equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific production equipment.
- The sample/s mentioned in this report were submitted/supplied/manufactured by the Client. TÜV SUD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SUD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SUD PSB or to the report or results furnished by TÜV SUD PSB in any advertisements or sales promotion.
- Unless otherwise stated, the tests were carried out in TÜV SUD PSB Pte Ltd, No 1 Science Park Drive Singapore 118221.
- The tests carried out by TÜV SUD PSB and this report are subject to TÜV SUD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SUD Group.

Effective 01 September 2020

Test Report No. dated 20 Oct 2020



Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SUD Group and the General Terms and Conditions of Business of TÜV SUD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

TESTED FOR:

TEST DATE:
22 Jul 2020 to 14 Aug 2020, 20 Oct 2020

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Nitrile Disposable Exam Gloves	Blue		XL	401	


Lot size as specified by client: 150,001 to 500,000 pieces

METHOD OF TEST:

- EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use
Part 3: Requirements and testing for biological evaluation
Clause 4.4 Powder-free gloves
Clause 4.6 Labelling


Yeo Poh Kwang Associate Engineer

Eng Din Yi Engineer Medical Health Services (NAM)



Phone: +65 6886 1333
Fax: +65 6761 8870
E-mail: enquiry@curagrip.com
http://www.curagrip.com.sg
Co Reg: 19022657R


Regional Head Office:
TUV SUD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221
TUV



INTERNATIONAL STANDARDS

EN 455:1-4

Test Report No.
dated 20 Oct 2020



RESULTS:

Sample: Nitrile Disposable Exam Gloves, Lot No.

Table 1: Results for EN 455-1:2020


Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4	Freedom from holes	Shall not leak	10	315	2	Passed

Table 2: Results for EN 455-2:2015 Clauses 4-5


Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	a) Length (mm)	≥ 240	13	249	Passed
	b) Width (mm)	For Size XL: ≥ 110	13	114	Passed
5	Strength a) Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	6.8	Passed
	b) Force at break after challenge testing (N) 7 days at 170±2°C	For nitrile examination gloves: ≥ 6.0	13	7.0	Passed

Table 3: Results for EN 455-2:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed



Test Report No.
dated 20 Oct 2020



RESULTS (cont'd):


Sample: Nitrile Disposable Exam Gloves, Lot No. Blue, Size XL

Table 4: Results for EN 455-3:2015 Clause 4.4

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.4	4.4	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	0.64 mg per glove	Passed

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex.	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: (Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses.	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free.	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: "CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions."	NA
d) for any medical glove containing nitrile rubber latex the product labelling shall not include:			
- any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein;	NA		
- any unjustified indication of the presence of allergens.	NA		
e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA		
		Inferred results	Passed



Test Report No.
dated 20 Oct 2020



REMARKS:

- Labelling requirements are assessed based on submitted packaging artwork by client on 20 Oct 2020.
- NA: Not applicable for the submitted sample.

Yeo Poh Kwang
Associate Engineer

Lee Chai Yi
Engineer
Medical Health Services (NAM)

APPENDIX:




Photo 1: Nitrile Disposable Exam Gloves, Lot No.



Photo 2: Packaging Artwork for Nitrile Disposable Exam Gloves, Lot No.





Test Report No.
dated 20 Oct 2020



Please note that this Report is issued under the following terms:

- This report applies to the sample of the specific production equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results shall not be used to indicate or imply that TÜV SUD PSB approves, recommends or endorses the manufacturer, supplier or user of such production equipment, or that TÜV SUD PSB in any way guarantees the later performance of the production equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific production equipment.
- The sample/s mentioned in this report shall be submitted/applied/manufactured by the Client. TÜV SUD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
- Nothing in this report shall be interpreted to mean that TÜV SUD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SUD PSB or to the report or results furnished by TÜV SUD PSB in any advertisements or sales promotion.
- Unless otherwise stated, the tests were carried out in TÜV SUD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.
- The tests carried out by TÜV SUD PSB and this report are subject to TÜV SUD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SUD Group.


Effective 01 September 2020


INTERNATIONAL STANDARDS

EN 374-1-2, EN 374-4-5, EN 21420

Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:	Auftrags-Nr.: Order no.:	Seite 1 von 28 Page 1 of 28	
Kunden-Referenz-Nr.: Client reference no.:	Auftragsdatum: Order date:	2022-01-28	
Auftraggeber: Client:			
Prüfgegenstand: Test item:	Schutzhandschuhe / Protective gloves		
Bezeichnung / Typ-Nr.: Identification / Type no.:	Nitrilhandschuhe Nitrile gloves		
Auftrags-Inhalt: Order content:	Produktüberwachung entsprechend Modul C2 der Verordnung (EU) 2016/425 product control according to Regulation (EU) 2016/425 of module C2		
Prüfungslage: Test specification:	EN ISO 374-1:2016 + A1:2018 Schutzhandschuhe gegen gefährliche Chemikalien und Mikroorganismen Protective gloves against dangerous chemicals and micro-organisms		
Wareneingangdatum: Date of sample receipt:	2022-04-22		
Prüfmuster-Nr.: Test sample no.:			
Prüfzeitraum: Testing period:	2022-05-25 – 2022-06-13		
Ort der Prüfung: Place of testing:	Prüfstelle für Textilien und PSA Köln		
Prüflaboratorium: Testing laboratory:	TÜV Rheinland LGA Products GmbH		
Prüfresultat: Test result:	Pass		
geprüft von: tested by:			
Datum: Date:	2022-06-13		
Stellung / Position: Sachverständige(r)/Expert	Sachverständige(r)/Expert		
Sonstiges / Other:	Prüfungen ausgewählter Parameter entsprechend o.g. Prüfungslage im Rahmen der Produktüberwachung gemäß Modul C2 der Verordnung (EU) 2016/425 Tests of selected parameters according to the above mentioned test basis within the scope of product monitoring in accordance with module C2 of Regulation (EU) 2016/425		
Zustand des Prüfgegenstandes bei Anlieferung: Condition of the test item at delivery:	Prüfmuster vollständig und unbeschädigt Test item complete and undamaged		
<p><small>* Legende: Pass = entspricht o.g. Prüfungsregeln / Fail = entspricht nicht o.g. Prüfungsregeln / NA = nicht anwendbar / Not applicable</small></p> <p><small>** Dieser Prüfbericht bezieht sich nur auf das o.g. Prüfmuster und darf ohne Genehmigung des Prüflagers nicht auszugswise vervielfältigt werden. Dieser Bericht berechtigt nicht zur Verwendung eines Prüflagers. This test report only relates to the above mentioned test sample. Without permission of the test center its test report is permitted to be duplicated in extracts. This test report does not entitle to carry any test report.</small></p>			
TÜV Rheinland LGA Products GmbH Tillystraße 2 D - 50433 Nürnberg Mail: service@de.tuv.com Web: www.tuv.com			



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:	Anmerkungen Remarks	Seite 2 von 28 Page 2 of 28	
<p>1 Alle eingesetzten Prüfmittel waren zum angegebenen Prüfzeitraum gemäß eines festgelegten Kalibrierungsprogramms unseres Prüflabors kalibriert. Sie entsprechen den in den Prüfprogrammen hinterlegten Anforderungen. Die Rückverfolgbarkeit der eingesetzten Prüfmittel ist durch die Einhaltung der Regelungen unseres Managementsystems gegeben. Detaillierte Informationen bezüglich Prüfbedingungen, Prüfequipment und Messunsicherheiten sind im Prüflabor vorhanden und können auf Wunsch bereitgestellt werden.</p> <p>The equipment used during the specified testing period was calibrated according to our test laboratory calibration program. The equipment fulfills the requirements included in the relevant standards. The traceability of the test equipment used is ensured by compliance with the regulations of our management system. Detailed information regarding test conditions, equipment and measurement uncertainty is available in the test laboratory and could be provided on request.</p> <p>2 Wie vertraglich vereinbart, wurde dieses Dokument nur digital unterzeichnet. Der TÜV Rheinland hat nicht überprüft, welche rechtlichen oder sonstigen diesbezüglichen Anforderungen für dieses Dokument gelten. Diese Überprüfung liegt in der Verantwortung des Benutzers dieses Dokuments. Auf Verlangen des Kunden kann der TÜV Rheinland die Gültigkeit der digitalen Signatur durch ein gesondertes Dokument bestätigen. Diese Anfrage ist an unseren Vertrieb zu richten. Eine Umweltgebühr für einen solchen zusätzlichen Service wird erhoben.</p> <p>As contractually agreed, this document has been signed digitally only. TÜV Rheinland has not verified and unable to verify which legal or other pertaining requirements are applicable for this document. Such verification is within the responsibility of the user of this document. Upon request by the client, TÜV Rheinland can confirm the validity of the digital signature by a separate document. Such request shall be addressed to our Sales department. An environmental fee for such additional service will be charged.</p> <p>3 Prüfklausel mit der Note * wurden an qualifizierte Unterunternehmer vergeben und sind unter der jeweiligen Prüfklausel des Berichts beschrieben. Abweichungen von Prüfspezifikationen) oder Kundenanforderungen sind in der jeweiligen Prüfklausel im Bericht aufgeführt.</p> <p>Test clauses with remark of * are subcontracted to qualified subcontractors and described under the respective test clause in the report. Deviations of testing specification(s) or customer requirements are in specific test clause in the report.</p> <p>4 Die Entscheidungsregel für Konformitätsfeststellungen in diesem Prüfbericht basiert auf der "Null-Grenzwert-Regel" und der "Einfachen Akzeptanz" gemäß ILAC G8:2019 und IEC Guide 115:2021, es sei denn, in der auf Seite 1 dieses Berichts genannten Norm ist etwas anderes festgelegt oder vom Kunden gewünscht. Dies bedeutet, dass die Messunsicherheit nicht berücksichtigt wird und daher auch nicht im Prüfbericht angegeben wird.</p> <p>The decision rule for statements of conformity in this test report is based on the "Zero Guard Band Rule" and "Simple Acceptance" in accordance with ILAC G8:2019 and IEC Guide 115:2021, unless otherwise specified in the applied standard mentioned on Page 1 of this report or requested by the customer. This means that measurement uncertainty is not taken in account and hence also not declared in the test report.</p> <p>5 Vorhersehbare Verwendung wurde betrachtet. Zuzett liegen für das/die Produkt/e weder Schutzklauselverfahren an, noch ist ein erhöhtes Unfallrisiko bekannt. Foreseeable use was considered. Currently neither a safeguard clause procedure has been introduced nor an increase in accidents known for this / these product (s).</p>			



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:	Anmerkungen Remarks	Seite 3 von 28 Page 3 of 28	
Überwachungshistorie / Surveillance history			
EU-Baumusterzertifikat-Nr.: EU Type Certificate No.:	i vom/ dated 2021-09-18		
Jahr / Year	geprüfte Abschnitt(e) / tested Clause(s)		
2019	EN 420, 5.1.2, 5.2; EN ISO 374-2, 7.1, 7.2, 7.3, EN ISO 374-4, 5.3, 5.3.4, 5.4, EN ISO 374-5, 5.1, 5.2		
2020	EN 420, 4.1, 4.2, 4.3, 5; EN ISO 374-2, 7.1, 7.2, 7.3, EN ISO 374-4, 5.3, 5.3.4, 5.4, 5.5, EN 420, 6, 7		
2021	EN ISO 374-2 Teil 7, 5.3; EN ISO 374-4, 5, EN ISO 374-5, 5.6, 6, 7; EN ISO 21420 Teil 4, Teil 7		
2022	EN 374-2 Teil 7, EN ISO 374-4 5.4, 5.5, EN ISO 374-5 5.1, 5.2, 5.4, Teil 6 und 7, EN 21420 Teil 4, 5 und 7		




Prüfbericht - Produkte Test Report - Products		TÜVRheinland®	
Prüfbericht-Nr.: Test report no.:	Produktbeschreibung Product description	Seite 4 von 28 Page 4 of 28	
1 Produktdetails Product details	5-Finger-Handschuh 5 finger gloves		
2 Artikel / Modell Article / Model	Nitrilhandschuhe Nitrile gloves		
3 Größe / Länge Size / Length	S (6,5), M (7,5), L(8,5), XL(9)		
4 Leistungsstufen Performance levels	Chemikale K NaOH 40%	Permeation: Klasse/ level 6	Degradation: -4,3 %
5 Verwendete Materialien Used materials	Nitril Nitrile Materialdicke 0,07 mm / wall thickness 0,07 mm		
6 Pflegekennzeichnung/ care instruction	n/a		
7 Mitgelieferte Dokumente / Prüfberichte Further applicable documents / test reports	*1 Prüfbericht Unsicherheiten / Test report innocuousness Bericht-Nr. / report nr.: vom/ or 2022-06-09		
8 Sonstiges Other	Test sample(s), as well sample information, description, product details and intended usage was provided by customer.		
9 Prüfmusterbereitstellung Test sample obtaining	<input type="checkbox"/> Sending by customer <input checked="" type="checkbox"/> Sampling by TÜV Rheinland Group Datum Musterziehung/sample picking date: 2022-03-31		
Verpackung/packaging			



INTERNATIONAL STANDARDS

EN 374-1-2, EN 374-4-5, EN 21420





Prüfbericht - Produkte
Test Report - Products

Prüfbericht-Nr.: Seite 5 von 28
Test report no.: Page 5 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
Der Originaltext wird nur auszugsweise wiedergegeben. Details sind dem Original-Dokument zu entnehmen. The original text is reproduced only in part. For details, be referred to the original document.			
1	Anwendungsbereich Scope		
2	Normative Verweisungen Normative references		
3	Begriffe Terms and definition		
4	Probennahme Sampling		
5	Leistungsanforderung Performance requirements		
5.1	Allgemeine Anforderungen General requirements		
5.2	Penetration Penetration		
5.3	Permeation Permeation		
5.4	Degradation Degradation		
5.5	Permeation Permeation		
5.6	Permeation Permeation		
5.7	Permeation Permeation		
5.8	Permeation Permeation		
5.9	Permeation Permeation		
5.10	Permeation Permeation		
5.11	Permeation Permeation		
5.12	Permeation Permeation		
5.13	Permeation Permeation		
5.14	Permeation Permeation		
5.15	Permeation Permeation		
5.16	Permeation Permeation		
5.17	Permeation Permeation		
5.18	Permeation Permeation		
5.19	Permeation Permeation		
5.20	Permeation Permeation		
5.21	Permeation Permeation		
5.22	Permeation Permeation		
5.23	Permeation Permeation		
5.24	Permeation Permeation		
5.25	Permeation Permeation		
5.26	Permeation Permeation		
5.27	Permeation Permeation		
5.28	Permeation Permeation		
5.29	Permeation Permeation		
5.30	Permeation Permeation		
5.31	Permeation Permeation		
5.32	Permeation Permeation		
5.33	Permeation Permeation		
5.34	Permeation Permeation		
5.35	Permeation Permeation		
5.36	Permeation Permeation		
5.37	Permeation Permeation		
5.38	Permeation Permeation		
5.39	Permeation Permeation		
5.40	Permeation Permeation		
5.41	Permeation Permeation		
5.42	Permeation Permeation		
5.43	Permeation Permeation		
5.44	Permeation Permeation		
5.45	Permeation Permeation		
5.46	Permeation Permeation		
5.47	Permeation Permeation		
5.48	Permeation Permeation		
5.49	Permeation Permeation		
5.50	Permeation Permeation		
5.51	Permeation Permeation		
5.52	Permeation Permeation		
5.53	Permeation Permeation		
5.54	Permeation Permeation		
5.55	Permeation Permeation		
5.56	Permeation Permeation		
5.57	Permeation Permeation		
5.58	Permeation Permeation		
5.59	Permeation Permeation		
5.60	Permeation Permeation		
5.61	Permeation Permeation		
5.62	Permeation Permeation		
5.63	Permeation Permeation		
5.64	Permeation Permeation		
5.65	Permeation Permeation		
5.66	Permeation Permeation		
5.67	Permeation Permeation		
5.68	Permeation Permeation		
5.69	Permeation Permeation		
5.70	Permeation Permeation		
5.71	Permeation Permeation		
5.72	Permeation Permeation		
5.73	Permeation Permeation		
5.74	Permeation Permeation		
5.75	Permeation Permeation		
5.76	Permeation Permeation		
5.77	Permeation Permeation		
5.78	Permeation Permeation		
5.79	Permeation Permeation		
5.80	Permeation Permeation		
5.81	Permeation Permeation		
5.82	Permeation Permeation		
5.83	Permeation Permeation		
5.84	Permeation Permeation		
5.85	Permeation Permeation		
5.86	Permeation Permeation		
5.87	Permeation Permeation		
5.88	Permeation Permeation		
5.89	Permeation Permeation		
5.90	Permeation Permeation		
5.91	Permeation Permeation		
5.92	Permeation Permeation		
5.93	Permeation Permeation		
5.94	Permeation Permeation		
5.95	Permeation Permeation		
5.96	Permeation Permeation		
5.97	Permeation Permeation		
5.98	Permeation Permeation		
5.99	Permeation Permeation		
5.100	Permeation Permeation		

TRF_EN ISO 374-1_2016 + A1:2018 (S/N=) - V1.0







Prüfbericht - Produkte
Test Report - Products

Prüfbericht-Nr.: Seite 6 von 28
Test report no.: Page 6 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
Carefully remove the glove from the wrapper, box or its packaging. Record the identity code, lot number, size and brand of samples. Visually examine for tears, rips and holes. If these are present, the gloves shall be reported as having failed.			
no tears, rips and holes are present			
7.2	Luft-Leck-Prüfung Air leak test		
4.1	Ein Handschuh wird in Wasser getaucht und sein Innenleben mit Luft aufgeblasen. Eine Undichtheit (Leck) wird als Strom aus Luftblasen sichtbar, der sich an der Oberfläche des Handschuhes bildet. A glove is immersed in water, and its interior is pressurized with air. A leak is detected by a stream of air bubbles from the surface of the glove.	Größe/size S keine/no Leakage M keine/no Leakage L keine/no Leakage XL keine/no Leakage	Luft-Leck-Prüfung/ Air leakage P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> N/T <input type="checkbox"/>
Tab. 1	Nennstärke der Handschuhe (e) nach Angaben des Herstellers mm Nominal glove thickness (e) mm As provided by the manufacturer.	Luftdruck (X) Air pressure (X) kPa	Verwendeter Luftdruck / air pressure used: 0,5 kPa
	e < 0,3	0,5	
	0,3 <= e < 0,5	2,0	
	0,5 <= e < 1,0	5,0	
	e > 1,0	6,0	
7.3	Wasser-Leck-Prüfung Water leak test		
4.2	Ein Handschuh wird mit Wasser gefüllt. Eine Undichtheit wird durch das Auftreten von Wassertröpfchen an der Außenseite des Handschuhes festgestellt. A glove is filled with water. A leak is detected by the appearance of water droplets on the outside of the glove.	Größe/size S keine/no Leakage M keine/no Leakage L keine/no Leakage XL keine/no Leakage	Wasser-Leck-Prüfung / Water leakage P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> N/T <input type="checkbox"/>
5.3	Degradation Degradation		P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> N/T <input type="checkbox"/>
5.4	Permeation Permeation		
5.4.1	Allgemeines General		
Für den Handschuh, der länger als 400 mm ist, und bei dem die Handinnenfläche und die Sohle unterschiedliche Leistungsstufen erreichen, muss für jede Chemikalie die geringere Leistungsstufe in der Kennzeichnung angegeben werden. Alle Ergebnisse sollen in der Benutzeranleitung angegeben werden. For the glove longer than 400 mm, where the palm and cuff achieve different performance levels, the lower performance level shall be claimed in the marking for each chemical. All the results should be reported in the user instruction.			

TRF_EN ISO 374-1_2016 + A1:2018 (S/N=) - V1.0







Prüfbericht - Produkte
Test Report - Products

Prüfbericht-Nr.: Seite 7 von 28
Test report no.: Page 7 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
Each combination of protective glove/ test chemical shall be classified according to Table 1, using the results as given in EN 16523-1:2015, 8.5.1.1 or 8.5.1.3 for the normalized breakthrough time.			
5.4.2	Typ A: Die Permeationsleistung muss mindestens Stufe 2 gegen wenigstens sechs Prüfchemikalien entsprechen, die in Tabelle 2 gelistet sind. Type A: The permeation performance shall be at least level 2 against a minimum of six test chemicals listed in Table 2.		
5.4.3	Typ B: Die Permeationsleistung muss mindestens Stufe 2 gegen wenigstens drei Prüfchemikalien entsprechen, die in Tabelle 2 gelistet sind. Type B: The permeation performance shall be at least level 2 against minimum of three test chemicals listed in Table 2.		
5.4.4	Typ C: Die Permeationsleistung muss mindestens Stufe 1 gegen wenigstens eine Prüfchemikalie entsprechen, die in Tabelle 2 gelistet ist. Type C: The permeation performance shall be at least level 1 against minimum of one test chemical listed in Table 2.		
Tab. 2	Kennbuchstabe Code Letter	Prüfchemikalie Chemical	CAS-RN CAS Number
	A	Methanol / Methanol	67-56-1
	B	Aceton / Acetone	67-64-1
	C	Acetonitril / Acetonitrile	75-05-8
	D	Dichlormethan / Dichloromethane	75-09-2
	E	Kohlendioxid / Carbon dioxide	75-15-0
	F	Toluol / Toluene	108-88-3
	G	Diethylamin / Diethylamine	109-89-7
	H	Tetrahydrofuran / Tetrahydrofuran	109-99-9
	I	Ethylacetat / Ethyl acetate	141-78-6
	J	n-Heptan / n-Heptane	142-92-5
	K	Natriumhydroxid 40 % / Sodium hydroxide 40 %	1310-73-2
	L	Schwefelsäure 96 % / Sulphuric acid 96 %	7804-93-9
	M	Salpetersäure 65 % / Nitric acid 65 %	7801-37-2
	N	Essigsäure 99 % / Acetic acid 99 %	64-19-7
	O	Ammoniakwasser 25 % / Ammonium hydroxide 25 %	1336-21-6
	P	Wasserstoffperoxid 30 % / Hydrogen peroxide 30 %	7722-84-1
	S	Flusssäure 40 % / Hydrofluoric acid 40 %	7664-39-3
	T	Formaldehyd 37 % / Formaldehyde 37 %	50-00-0
Leistungsstufen gegen Permeation Permeation performance levels			
Tab. 1	Gemessene Durchbruchzeit / Measured breakthrough time [min]	Schutzindex / Protection performance level	Prüfchemikalie / Chemical
	> 10	Klasse / class 1	Natriumhydroxid 40 % / Sodium hydroxide 40 %
	> 30	Klasse / class 2	
	> 60	Klasse / class 3	
	> 120	Klasse / class 4	
	> 240	Klasse / class 5	
	> 480	Klasse / class 6	
Die Prüfchemikalie(n) muss (müssen) aus der Liste der Prüfchemikalien in Tabelle 2 genommen werden.			

TRF_EN ISO 374-1_2016 + A1:2018 (S/N=) - V1.0






Prüfbericht - Produkte
Test Report - Products

Prüfbericht-Nr.: Seite 8 von 28
Test report no.: Page 8 of 28

Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result			
Abhängig von der Anwendung der Handschuhe könnten andere Prüfchemikalien verwendet werden. The test chemical(s) shall be taken from the list of test chemicals in Table 2. Other test chemicals could be used depending on the application of the gloves.						
5.5	Anforderungen an Handschuh-Typen A, B und C Requirements for gloves types A, B and C					
Anforderungen an verschiedene Schutztypen von Handschuhen Requirements for different protection types of gloves						
Tab. 3						
	Typ A / Type A	5.1	5.2	5.4.2	5.4.3	5.4.4
	Typ B / Type B	X	X	X	X	
	Typ C / Type C	X	X	X	X	X
	X = erforderlich / required					
EN ISO 374-5	Teil 5: Terminologie und Leistungsanforderungen für Risiken durch Mikroorganismen Part 5: Terminology and performance requirements for micro-organisms risks					
5	Leistungsanforderung Performance requirement					
5.1	Allgemeine Anforderungen General requirements					
Schutzhandschuhe gegen Mikroorganismen sollen der EN ISO 21420:2020, Absatz 4, Abs. 5 und Abs. 7 entsprechen. Protective gloves against micro-organisms risks shall comply with the requirements given in EN ISO 21420:2020, Clause 4, Clause 5 and Clause 7.						
5.2	Penetration Penetration					
Schutzhandschuhe gegen Viren, Bakterien und Pilze dürfen bei der Prüfung nach EN 374-2:2019, 7.2 und 7.3 nicht undicht werden. Protective gloves against virus, bacteria and fungi shall not leak when tested according to EN 374-2:2019, 7.2 and 7.3.						
5.3	Schutz vor Viren Protection against viruses					
Schutzhandschuhe gegen Viren sind nach ISO 16604 Verfahren B zu testen und dürfen im Testkitter keinen nachweisbaren Transfer (c1 PFU/ml) des Phi-X174-Bakteriophagen aufweisen. Protective gloves against virus shall be tested according to ISO 16604 Procedure B and shall exhibit no detectable transfer (c1 PFU/ml) of the Phi-X174 bacteriophage in the assay titre.						
	nicht anwendbar, da nicht ausgetestet					
	not applicable, because not marked					


TRF_EN ISO 374-1_2016 + A1:2018 (S/N=) - V1.0




INTERNATIONAL STANDARDS

EN 374-1-2, EN 374-4-5, EN 21420

Prüfbericht - Produkte Test Report - Products		TÜVRheinland®													
Prüfbericht-Nr.: Test report no.:		Seite 9 von 29 Page 9 of 29													
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result												
5.4	Anforderungen an verschiedene Schutzarten von Handschuhen Requirements for different protection types of gloves Die Anforderungen sind in der Tabelle 1 aufgeführt. The requirements are mentioned in the Table 1.														
Tab. 1	<table border="1"> <thead> <tr> <th></th> <th>5.1</th> <th>5.2</th> <th>5.3</th> </tr> </thead> <tbody> <tr> <td>Handschuh gegen Bakterien und Pilze (Glove protecting against bacteria and fungi)</td> <td>X</td> <td>X</td> <td></td> </tr> <tr> <td>Handschuh gegen Viren, Bakterien und Pilze (Glove protecting against virus, bacteria and fungi)</td> <td>X</td> <td>X</td> <td>X</td> </tr> </tbody> </table> <p>X = erforderlich / required</p>		5.1	5.2	5.3	Handschuh gegen Bakterien und Pilze (Glove protecting against bacteria and fungi)	X	X		Handschuh gegen Viren, Bakterien und Pilze (Glove protecting against virus, bacteria and fungi)	X	X	X		
	5.1	5.2	5.3												
Handschuh gegen Bakterien und Pilze (Glove protecting against bacteria and fungi)	X	X													
Handschuh gegen Viren, Bakterien und Pilze (Glove protecting against virus, bacteria and fungi)	X	X	X												
6	Kennzeichnung Marking Die Kennzeichnung von Schutzhandschuhen gegen gefährliche Chemikalien muss mit der Anforderung an Schutzhandschuhe in EN ISO 21420:2020 und mit folgenden Punkten übereinstimmen. All information shall be precise and comprehensive, and provided at least in the official language(s) of the country of destination.														
6.1	Kennzeichnung von Handschuhen des Typ A Marking of Type A gloves Bild / Fig. 2 Für Schutzhandschuhe, die die in 5.5 angegebenen Typ-A-Anforderungen erfüllen, ist das Piktogramm in Bild 2 mit Verweisung auf diesen Teil von ISO 374-1 zu verwenden. Die sechs geprüften Chemikalien müssen durch ihren Kennbuchstaben identifiziert werden, die unterhalb des Piktogramms angegeben werden müssen, wie in Bild 2 dargestellt. Wurden weitere Chemikalien geprüft, die nicht in der Liste angegeben sind, müssen die Informationen über die Leistungsstufen in der Benutzeranleitung zur Verfügung gestellt werden. For protective gloves complying with the type A requirements stated in 5.5, the pictograms in Figure 2 shall be used with reference to this part of ISO 374-1. The six tested chemicals shall be identified by their code letter which shall be marked under the pictogram as shown in Figure 2. If other chemicals not present in the list have been tested, information about the performance levels shall be provided in the user instructions.	nicht anwendbar, da Typ C	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>												
6.2	Kennzeichnung von Handschuhen des Typ B Marking of Type B gloves Bild / Fig. 3 Für Schutzhandschuhe, die die in 5.5 angegebenen Typ-B-Anforderungen erfüllen, ist das Piktogramm in Bild 3 mit Verweisung auf diesen Teil von ISO 374-1 zu verwenden. Die drei geprüften Chemikalien müssen durch ihren Kennbuchstaben identifiziert werden, die unterhalb des	nicht anwendbar, da Typ C	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>												



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®		
Prüfbericht-Nr.: Test report no.:		Seite 10 von 28 Page 10 of 28		
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result	
5.4	Piktogramm angegeben werden müssen, wie in Bild 3 dargestellt. Wurden weitere Chemikalien geprüft, die nicht in der Liste angegeben sind, müssen die Informationen über die Leistungsstufen in der Benutzeranleitung zur Verfügung gestellt werden. For protective gloves complying with the type B requirements stated in 5.5, the pictograms in Figure 3 shall be used with reference to this part of ISO 374-1. The three tested chemicals shall be identified by their code letter which shall be marked under the pictogram as shown in Figure 3. If other chemicals not present in the list have been tested, information about the performance levels shall be provided in the user instructions.			not applicable, because type C
6.3	Kennzeichnung von Handschuhen des Typ C Marking of Type C gloves Bild / Fig. 4 Für Schutzhandschuhe, die die in 5.5 angegebenen Typ-C-Anforderungen erfüllen, ist das Piktogramm in Bild 4 mit Verweisung auf diesen Teil von ISO 374-1 zu verwenden. Die getestete Chemikalie muss in der Gebrauchsanweisung mit Angaben zu ihrer Leistungsstufe angegeben werden. Wurden weitere Chemikalien geprüft, die nicht in der Liste angegeben sind, müssen die Informationen über die Leistungsstufen in der Benutzeranleitung zur Verfügung gestellt werden. For protective gloves complying with the type C requirements stated in 5.5, the pictogram in Figure 4 shall be used and the reference to this part of ISO 374-1. The tested chemical shall be given in the user instructions with information about its performance levels. If other chemicals not present in the list have been tested, information about the performance levels shall be provided in the user instructions.			
6.2	Kennzeichnung Mikroorganismen Marking microorganisms Schutzhandschuhe, die vor Mikroorganismen schützen, müssen den Anforderungen der EN ISO 21420:2020 Absatz 4, 5 und 7 entsprechen. Sie dürfen keine Beschädigung bei der Prüfung auf Wasser-Leck und Luft-Leck gemäß EN 374-2:2019 aufweisen. Protective gloves against micro-organism risks shall comply with the requirements given in EN ISO 21420:2020, Clause 4, Clause 5 and Clause 7.	gegeben, Handschuhe entsprechen Anforderungen des EN 21420:2020 Absatz 4.5 und 7, sowie weisen keine Beschädigung bei der Prüfung auf Wasser-Leck und Luft-Leck auf	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>	



Prüfbericht - Produkte Test Report - Products		TÜVRheinland®		
Prüfbericht-Nr.: Test report no.:		Seite 11 von 29 Page 11 of 29		
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result	
	Protective gloves against virus, bacteria and fungi shall not leak when tested according to EN 374-2:2019, 7.2 Air leak test and 7.3 water leak test. ISO 374-5:2016	damage during testing of water leak and air leak		
6.3	Kennzeichnung von Handschuhen, die vor Viren, Bakterien und Pilze schützen Marking of gloves protecting against viruses, bacteria and fungi Schutzhandschuhe die vor Viren schützen, müssen den Anforderungen aus EN 374-5 6.2 entsprechen und dürfen gemäß ISO 16604 Verfahren B kein nachweisbarer Transfer (<1 PFU/ml) des Phi-X174 Bakteriophagen bei der Titration aufweisen. Protective gloves against virus have to comply with the requirements of EN 374-5 6.2 and shall be tested according to ISO 16604 Procedure B and shall exhibit no detectable transfer (<1 PFU/ml) of the Phi-X174 bacteriophage in the assay titre. ISO 374-5:2016	nicht anwendbar, da nicht ausgetestet	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>	
7	Information des Herstellers Information supplied by the manufacturer EN 374-1 Die Informationen des Herstellers müssen in Übereinstimmung mit den Anforderungen an die Informationen stehen, die in EN ISO 21420 festgelegt sind.	gegeben	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>	




Prüfbericht - Produkte Test Report - Products		TÜVRheinland®		
Prüfbericht-Nr.: Test report no.:		Seite 12 von 28 Page 12 of 28		
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result	
	Sie müssen außerdem die Ergebnisse von 5.2, 5.3, 5.4 (Perforation, Degradation, Permeation) enthalten. Die Liste sämtlicher Chemikalien, auf die die Schutzhandschuhe geprüft wurden und die Leistungsstufen, die bei der Permeationprüfung erreicht wurden. Folgende Warnhinweise müssen in der Benutzeranleitung hinzugefügt werden: „Diese Information macht keine Angaben zur tatsächlichen Schutzdauer am Arbeitsplatz und zur Unterscheidung von Gemischen und reinen Chemikalien.“ „Der Widerstand gegen Chemikalien wurde unter Laborbedingungen an Proben beurteilt, die lediglich von der Handinnenfläche entnommen wurden (ausgenommen ist der Fall, bei dem der Handschuh 400 mm oder länger ist - in diesem Fall wird ebenfalls die Stulpe getestet) und bezieht sich ausschließlich auf die geprüften Chemikalien. Er kann anders sein, wenn die Chemikalie in einem Gemisch verwendet wird.“ „Es wird eine Überprüfung empfohlen, ob die Handschuhe für die vorgesehene Verwendung geeignet sind, da die Bedingungen am Arbeitsplatz in Abhängigkeit von Temperatur, Abrieb und Degradation von denen der Typprüfung abweichen können.“ „Wurden Schutzhandschuhe bereits verwendet, können sie aufgrund von Veränderungen ihrer physikalischen Eigenschaften geringeren Widerstand gegen gefährliche Chemikalien bieten. Durch die Berührung mit Chemikalien verursachte Degradation, Bewegungen, Fadenzühen, Reibung usw. kann die tatsächliche Anwendungsdauer wesentlich reduziert werden. Bei aggressiven Chemikalien kann die Degradation der wichtigste Faktor sein, der bei der Auswahl von gegen Chemikalien beständigen Handschuhen zu berücksichtigen ist.“ „Vor der Anwendung sind die Handschuhe auf jegliche Fehler oder Mängel zu überprüfen.“ Bei Handschuhen, die mehrfach verwendet werden können, muss der Hersteller die relevanten Anleitungen für die Dekontamination angeben. Ist keine Information zur Dekontamination vorhanden, sind die Handschuhe nur für die einmalige Verwendung vorgesehen und folgende Warnhinweise ist hinzu zu fügen: „Nur für die einmalige Verwendung bestimmt.“ The information supplied by the manufacturer shall be in accordance with the requirements for information as defined in EN ISO 21420.	Das Schutzlevel ist mit Level 6 angegeben, (K) NaOH 40%, Degradation bei -4,3% gegeben auf zusätzlichem Einleger gegeben in englisch gegeben in englisch gegeben in englisch gegeben in englisch gegeben in englisch n/a gegeben in englisch		
	The protection level is given as level 6, (K) NaOH 40%, Degradation at -4,3%			




INTERNATIONAL STANDARDS


EN 374-1-2, EN 374-4-5, EN 21420

			
Prüfbericht - Produkte			
Test Report - Products			
Prüfbericht-Nr.:		Seite 13 von 28	
Test report no.:		Page 13 of 28	
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	<p>EN ISO 374-1:2016 + A1:2018</p> <p>It shall also include the results of 5.2, 5.3, 5.4 (Penetration, Degradation, Permeation), the list of all the chemicals to which the protective gloves have been tested and the performance levels obtained in permeation testing.</p> <p>The following warnings shall be added in the user instructions: "This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals." "The chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in cases where the glove is equal to or over 400 mm - where the cuff is tested also) and relates only to the chemical tested. It can be different if the chemical used in a mixture."</p> <p>"It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on temperature, abrasion and degradation."</p> <p>"When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves"</p> <p>"Before usage, inspect the gloves for any defect or imperfections."</p> <p>For reusable gloves, the manufacturer shall provide the relevant instructions for decontamination. If there is no information about decontamination, then it is intended for single use and the following warning shall be added: "For single use only".</p>	<p>given on additional insert</p> <p>given</p> <p>given</p> <p>given</p> <p>given</p> <p>given</p> <p>n/a</p> <p>given</p>	<p>EN ISO 374-1:2016 + A1:2018</p> <p>gegeben mit:</p> <p>gegeben in englisch</p>
EN ISO 374-5	<p>Schutzhandschuhe, die gekennzeichnet sind Schutz gegen Micro-Organismen zu tragen und den Anforderungen von 5.4 entsprechen, ist dies in der Informationsbrochure anzugeben.</p> <p>Die folgende Warnung sollte hinzugefügt werden, dass diese Information nicht die tatsächliche Leistung am Arbeitsplatz widerspiegelt: "Die Penetration wurde</p>	<p>gegeben mit:</p> <p>gegeben in englisch</p>	




			
Prüfbericht - Produkte			
Test Report - Products			
Prüfbericht-Nr.:		Seite 14 von 28	
Test report no.:		Page 14 of 28	
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	<p>EN ISO 374-1:2016 + A1:2018</p> <p>unter Laborbedingung bewertet und bezieht sich nur auf die geprüften Proben!</p> <p>Falls nicht gegen Viren geprüft: "Nicht gegen Viren getestet!"</p> <p>For protective gloves that are marked offering protection against micro-organisms and complying with the requirements in 5.4, this shall be stated in the user instructions.</p> <p>The following warning shall be added that this information does not reflect the actual performance in the workplace. The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen. If not tested against viruses, the following warning shall be added: "Not tested against viruses".</p>	<p>gegeben in englisch</p> <p>given by:</p> <p>given</p> <p>given</p>	<p>EN ISO 374-1:2016 + A1:2018</p> <p>gegeben mit:</p> <p>gegeben in englisch</p>



				
Prüfbericht - Produkte				
Test Report - Products				
Prüfbericht-Nr.:		Seite 15 von 28		
Test report no.:		Page 15 of 28		
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result	
	<p>EN ISO 21420:2020 - Schutzhandschuhe - Allgemeine Anforderungen und Prüfverfahren</p> <p>EN ISO 21420:2020 - Protective gloves - General requirements and test methods</p> <p>1 Anwendungsbereich Scope</p> <p>2 Normative Verweisungen Normative references</p> <p>3 Begriffe Terms and definitions</p> <p>4 Anforderungen Requirements</p> <p>4.1 Gestaltungsgrundsätze und Handschuhkonfektionierung - Allgemeines Glove design and construction - General</p> <p>Der Schutzhandschuh muss so konzipiert und hergestellt sein, dass der Träger unter den vorhersehbaren Einsatzbedingungen die Tätigkeit so normal wie möglich ausführen kann und dabei über einen angemessenen Schutz verfügt.</p> <p>Dieses Dokument muss gemeinsam mit den zutreffenden spezifischen Normen angewendet werden, um diese Angemessenheit zu verifizieren.</p> <p>Wenn es von der entsprechenden spezifischen Norm (z. B. ISO 16073:2011, 5.7.3) gefordert wird, muss der Handschuh so gestaltet werden, dass die für das Anziehen und Ausziehen des Handschuhs benötigte Zeit minimiert wird.</p> <p>Wiederverwendbare mehrlagige Handschuhe müssen sich ohne Lösen der einzelnen Lagen der Finger voneinander ausziehen lassen. Wenn die Handschuhkonfektionierung Nähte mit einschließt, muss das Material und die Festigkeit der Nähte so beschaffen sein, dass die gesamte Leistung des Handschuhs nicht wesentlich herabgesetzt wird, wie von den entsprechenden spezifischen Normen gefordert.</p> <p>The protective glove shall be designed and manufactured so that in the foreseeable conditions of use, the wearer can perform the activity as normally as possible with an appropriate protection.</p> <p>This document along with the appropriate specific standards shall be used to verify this adequacy.</p> <p>If required in the relevant specific standard (for example ISO 16073:2011, 5.7.3), the glove shall be designed to minimize the donning and doffing time.</p> <p>For reusable multilayer gloves, the gloves shall be able to be doffed without separation of the layers of the fingers. When the glove construction includes seams, the material and strength of the seams shall be such that</p>	<p>gegeben</p> <p>gegeben mit EN 374</p> <p>n/a</p> <p>n/a</p> <p>given</p> <p>given with EN 374</p> <p>n/a</p> <p>n/a</p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>	





				
Prüfbericht - Produkte				
Test Report - Products				
Prüfbericht-Nr.:		Seite 16 von 28		
Test report no.:		Page 16 of 28		
Absatz Clause	Anforderungen - Prüfungen / Requirements - Tests	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result	
	<p>EN ISO 374-1:2016 + A1:2018</p> <p>the overall performance of the glove is not significantly decreased as required in the relevant specific standards.</p> <p>4.2 Unschädlichkeit von Schutzhandschuhen Innocuousness of protective gloves</p> <p>Schutzhandschuhe dürfen sich nicht nachteilig auf die Gesundheit und Hygiene des Benutzers auswirken. Die Materialien sollten unter den vorhersehbaren Bedingungen der üblichen Anwendung keine Stoffe freisetzen, die allgemein als toxisch, karzinogen, mutagen, allergen, reproduktionstoxisch, ätzend, sensibilisierend oder reizend bekannt sind.</p> <p>Protective gloves shall not adversely affect the health or hygiene of the user. The materials should not, in the foreseeable conditions of normal use, release substances generally known to be toxic, carcinogenic, mutagenic, allergenic, toxic to reproduction, corrosive, sensitizing or irritating.</p> <p>a Bestimmung des Chrom(VI)-Gehaltes Determination of chromium (VI) content</p> <p>< 30 mg/kg</p> <p>nach / according to: nach ISO 17075-1</p> <p>Enthält der Handschuh verschiedene Arten von Leder, muss jede Lederart, unabhängig davon, ob sie mit der Haut in Berührung kommt oder nicht, separat geprüft werden und die vorgenannte Anforderung erfüllen.</p> <p>If the glove includes different types of leather, whether in contact with the skin or not, each leather type shall be tested separately and comply with the above requirement.</p> <p>b Nickelabgabe / release of nickel</p> <p>Alle metallischen Materialien, mit denen die Haut längere Zeit in Berührung kommen könnte (z. B. Knöpfe, Zubehörfteile), müssen eine Nickelabgabemenge von < 0,5 µg/cm² je Woche aufweisen.</p> <p>All metallic materials which could come into prolonged contact with the skin (for example studs, fittings) shall have a release of nickel of less than 0.5 µg/cm² per week.</p> <p>nach / according to: EN 1811-A1:2015</p> <p>1.1 Bestimmung des pH-Wertes Determination of pH-value</p> <p>Alle Handschuhmaterialien müssen folgenden pH-Wert aufweisen: All glove materials shall have a pH value of: > 3,5 und/ oder < 9,5.</p> <p>nach / according to: ISO 4045 für Leder/ for leather ISO 3071 andere Materialien/ other materials</p>	<p>gegeben durch Herstellerklärung und Prüfung aller relevanten Parameter nach Abs. 4.2 a-f</p> <p>given by manufacturer's declaration and testing of all relevant parameter acc. clause 4.2 a-f</p> <p>n/a=kein Leder/no leather</p> <p>n/a= keine metallischen Materialien/no metallic materials</p> <p>9,3</p>	<p>P <input checked="" type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p> <p>P <input type="checkbox"/></p> <p>F <input type="checkbox"/></p> <p>N/A <input type="checkbox"/></p> <p>NT <input type="checkbox"/></p>	













INTERNATIONAL STANDARDS




EN 374-1-2, EN 374-4-5, EN 21420

			
Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 21 von 29 Page 21 of 29	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	- If this property is claimed for a leather glove, it shall have a water vapour transmission of at least 5 mg/(cm ² ·h) when tested according to 6.3.1. - If this property is claimed for a textile glove, it shall have a water vapour resistance less than or equal to 30 m ² ·Pa·W when tested according to 6.3.2.		
5.3.2	Wasserdampfaufnahme Water vapour absorption Wenn die schützenden Eigenschaften des Handschuhs die Wasserdampfdurchlässigkeit verhindern oder ausschließen, müssen Handschuhe, falls praktisch durchführbar, so konzipiert sein, dass die Schweißaufnahme so stark wie möglich reduziert wird. Wenn diese Eigenschaft für einen Ledershandschuh angegeben ist, muss er bei Prüfung nach 6.4.2 eine Wasserdampfaufnahmefähigkeit von ≥ 8 mg/(cm ² ·h) aufweisen. Where the protection characteristics of the glove inhibit or exclude water vapour transmission, when practicable, the gloves shall be designed to reduce the perspiration absorption as much as possible. If this property is claimed for a leather glove, it shall have a water vapour absorption of at least 8 mg/(cm ² for h when tested according to 6.4.2.	nicht anwendbar, auf Grund der Schutzwirkung gegen gefährliche Chemikalien und Mikroorganismen not applicable, because of the protection against dangerous chemicals and micro-organisms	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
6	Prüfverfahren Test procedures		
7	Kennzeichnung und Information Marking and information		
7.1	Allgemeines General		
7.1.1	Alle Informationen müssen präzise und nachvollziehbar sein. All information shall be precise and comprehensive.	gegeben given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.2	Kennzeichnung Marking		
7.2.1	Handschuhkennzeichnung Glove marking		
7.2.1.1	Jeder Schutzhandschuh muss mit folgenden Angaben gekennzeichnet sein: a. Name, Handelsmarke oder andere Erkennungsmerkmale des Herstellers oder des bevollmächtigten Repräsentanten des Herstellers; b. Handschuhbezeichnung (Handelsname oder Code, die dem Benutzer die eindeutige Identifizierung des Produkts innerhalb des Sortiments des Herstellers bzw. des bevollmächtigten Repräsentanten ermöglichen);	Kennzeichnung gegeben auf der Verpackung	P <input checked="" type="checkbox"/> F <input checked="" type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>







			
Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 22 von 29 Page 22 of 29	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	c. Größenbezeichnung; d. wenn der Handschuh einer oder mehreren spezifischen Norm(en) entspricht (siehe Literaturliste), muss/müssen das/die graphische(n) Symbol(e) den Angaben in Anhang C entsprechen. Jedes graphische Symbol muss zusammen mit der Verweisung auf die anwendbare spezifische Norm und den Leistungsstufen angegeben werden (siehe 7.3.5), die stets in derselben feststehenden Reihenfolge angegeben werden müssen, die in der entsprechenden Norm festgelegt ist. e. Herstellungsdatum, zumindest Monat und Jahr (z. B. 2016/11), oder andere Mittel, mit denen die Rückverfolgbarkeit der Chargen sichergestellt wird; f. wenn anwendbar, das Ablaufdatum, zumindest Monat und Jahr (z. B. 2016/11), hinter dem graphischen Symbol der Sanduhr, wie in Anhang C dargestellt. Each protective glove shall be marked with the following information: a. Name, trade mark or other means of identification of the manufacturer or the manufacturer's authorized representative; b. Glove designation (commercial name or code allowing the user to identify clearly the product within the manufacturer's authorized representative's range); c. Size designation; d. Where the glove conforms to one or more specific standards (see Bibliography), the pictogram (s) shall be accompanied by the reference of the applicable specific standard and performance levels (see 7.3.5), which shall always be in the same fixed sequence as defined in the corresponding standard; e. Date of manufacturing, at least the month and year (for example 2016/11), or any mean ensuring the manufacturing batch traceability; g. If applicable, the obsolescence date, at least the month and year (for example 2016/11), behind the hour glass pictogram as shown in Annex C.	c. gegeben mit z.B. Gr. M (7.5) d. gegeben mit:  e. gegeben mit:  f. gegeben mit:  g. gegeben mit:  c. given by e.g. size M (7.5) d. given by:  e. given by:  f. given by:  g. given by: 	
7.2.1.2	Die Kennzeichnung muss über die gesamte vorhersehbare Lebensdauer deutlich sichtbar und leicht angebracht sein. Kennzeichnungen oder Aufsticker, die zu Verwechslungen mit den obigen Kennzeichnungen führen könnten, dürfen nicht am Handschuh angebracht werden. The marking shall be affixed so as to be visible and legible throughout the foreseeable useful life of the glove. Marks or inscriptions which could be confused with the above marks shall not be affixed to the glove.	gegeben an der Verpackung, Kennzeichnung am Handschuh nicht möglich given on packaging, marking on glove not possible	P <input checked="" type="checkbox"/> F <input checked="" type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>



			
Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 23 von 29 Page 23 of 29	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.2.1.3	Sofern die Kennzeichnung auf dem Handschuh aufgrund der Produktgestaltung nicht möglich ist, ist sie auf der Verpackung oder einem dem Handschuh beiliegenden Dokument anzubringen. If marking on the glove is not possible given the characteristics of the product, the marking shall be affixed to the packaging or any document supplied with the glove.	gegeben, auf der Verpackung given, on packaging	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.2.1.4	Ein graphisches Symbol darf nur angegeben werden, wenn der Handschuh die Mindestanforderungen der entsprechenden spezifischen Norm erfüllt. A pictogram shall only be used when the glove meets at least the minimum requirement of the relevant specific standard.	gegeben given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.2.2	Kennzeichnung der Verpackung Marking of packaging		
	Jede Verpackung, die die Handschuhe unmittelbar enthält, muss eindeutig mit folgenden Angaben gekennzeichnet sein: a. Name und vollständige Anschrift des Herstellers oder des bevollmächtigten Repräsentanten des Herstellers; b. die in 7.2.1.1 b) und c) geforderten Informationen; c. Hinweis, wo die Informationen nach 7.3 erhalten werden können; d. wenn es sich um einen einfachen Handschuh handelt, der dem Benutzer nur gegen Gefahren wie die in Anhang A aufgeführten schützen soll, müssen die Worte „Nur für minimale Risiken“ oder eine ähnliche Formulierung aufgedruckt werden; e. das/die der spezifischen Norm entsprechende(n) graphische(n) Symbol(e), siehe Anhang C, wenn der Handschuh dieser spezifischen Norm entspricht (siehe Literaturliste); jedem graphischen Symbol müssen eine Verweisung auf die entsprechende Norm sowie die Leistungsstufen hinzugefügt werden, und zwar immer in derselben Reihenfolge, wie sie in der zutreffenden spezifischen Norm festgelegt sind. Wenn zusätzliche graphische Symbole genutzt werden, müssen sie in den Informationen des Herstellers erläutert werden (7.3); f. sofern zutreffend, eine nach 7.3.6 geforderte Angabe; g. sofern anwendbar, das Ablaufdatum, zumindest Monat und Jahr (z. B. 2016/11), hinter dem graphischen Symbol der Sanduhr, wie in Anhang C dargestellt.	a. gegeben B. Gr. M (7.5) c. gegeben mit Piktogramm d. n/a da PSA Kat. III e. gegeben mit:  f. n/a g. gegeben mit: 	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>



			
Prüfbericht - Produkte Test Report - Products			
Prüfbericht-Nr.: Test report no.:		Seite 24 von 29 Page 24 of 29	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
	Each packaging enclosure that immediately contains the gloves shall be clearly marked with the following: a. Name and full address of the manufacturer or the manufacturer's authorized representative; b. The information required in 7.2.1.1 b) and c); c. Reference to where the information required in 7.3 may be obtained; d. Where the glove is of simple design intended to protect the wearer against only those hazards listed in Annex A, the words "For minimal risks only" or an equivalent expression shall be printed; e. The pictogram(s) appropriate to the specific standard, see Annex C, when the glove conforms to this specific standard (see Bibliography). Each pictogram shall be accompanied by the performance levels, which shall always be in the same fixed sequence as defined in the relevant specific standard, and the reference to the applicable standard. If additional pictograms are used, they shall be explained in the information supplied by the manufacturer (7.3); f. Where applicable, information required in 7.3.6; g. If applicable, the obsolescence date, at least the month and year (for example 2016/11), behind the hour glass pictogram as shown in Annex C.	a. given B. Gr. M (7.5) c. given by pictogram d. n/a - because PPE Cat. III e. given by:  f. n/a g. given by: 	
7.3	Informationen des Herstellers Information supplied by the manufacturer		
	Wenn der Schutzhandschuh auf den Markt gebracht wird, müssen folgende Mindestinformationen bereitgestellt und verfügbar gehalten werden: The following minimum information shall be supplied when the protective glove is placed on the market and shall be maintained available:	gegeben given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.1	Name und vollständige Anschrift des Herstellers oder seines bevollmächtigten Repräsentanten. Name and full address of the manufacturer or authorized representative.	gegeben/given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.2	Handschuhbezeichnung nach 7.2.1.1 b) ; Glove designation as per 7.2.1.1 b).	gegeben/given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.3	Information zu dem verfügbaren Größenbereich und, sofern zutreffend, die nach 5.1 geforderten Informationen. Information on the available size range and where applicable, information required in 5.1.	gegeben mit S (6.5) bis XL (8) given by S (6.5) to XL (8)	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>



INTERNATIONAL STANDARDS

EN 374-1-2, EN 374-4-5, EN 21420

TÜVRheinland® Prüferbericht - Produkte Test Report - Products			
Prüferbericht-Nr.: Test report no.:		Seite 25 von 29 Page 25 of 29	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.4	Der bestimmungsgemäße Gebrauch des Handschuhs und eine Verweisung auf die entsprechende(n) spezifische(n) Norm(en) und das Jahr der Veröffentlichung (siehe Literaturhinweise). The intended use of the glove and reference to the relevant specific standard(s) and publication year (see Bibliography).	gegeben in englisch given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.5	Wenn nach 7.2.1.1 d) und 7.2.2 e) zutreffend, grafische Symbole, die die Gefahrenkategorien angeben, gegebenenfalls gefolgt von den Leistungsstufen. 0: gibt an, dass der Handschuh unter die Mindestleistungsstufe für eine bestimmte einzelne Gefahr fällt; X: gibt an, dass der Handschuh nicht geprüft wurde oder das Prüfverfahren für die Handschuhkonfektionierung oder das Handschuhmaterial ungeeignet scheint. Weiterhin sind grundsätzliche Erklärungen beizufügen, um das Verstehen der relevanten Leistungsstufen zu unterstützen. Die Normen, auf die sie sich beziehen, sind anzugeben. Die Gründe für die Angabe „X“ müssen erklärt werden. Die Leistungsstufen müssen in derselben Reihenfolge wie in der entsprechenden spezifischen Norm angegeben werden. Sie dürfen an einer beliebigen Stelle in der Nähe des grafischen Symbols angegeben werden, vorausgesetzt, sie stehen dazu in einem deutlichen Bezug. Where applicable as per 7.2.1.1 d) and 7.2.2 e), pictogram(s) indicating categories of hazard followed as applicable by the performance levels. 0: indicates that the glove falls below the minimum performance level for the given individual hazard. X: indicates that the glove has not been tested or the test method appears not to be suitable for the glove design or material. Furthermore, a basic explanation shall be given to assist comprehension of the relevant performance levels, and the standard(s) to which they refer shall be indicated. The reason(s) to use "X" shall be explained. Performance level shall be in the same order as given within the relevant specific standard. They may be positioned anywhere next to the pictogram provided that they are in clear relation with it.	n/a- EN 374 Performance level: 0: <input checked="" type="checkbox"/> 1: <input checked="" type="checkbox"/> 2: <input checked="" type="checkbox"/>	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>



TÜVRheinland® Prüferbericht - Produkte Test Report - Products			
Prüferbericht-Nr.: Test report no.:		Seite 26 von 28 Page 26 of 28	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.6	Wenn der Schutz nur auf einen Teil der Hand beschränkt ist, ist dies zu erwähnen. When protection is limited to part of the hand only, this shall be mentioned.	n/a- da Schutz nicht nur auf einen Teil der Hand beschränkt ist	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.3.7	Sofern zutreffend, müssen Warnungen hinsichtlich möglicherweise eintretender Probleme oder Einschränkungen bei der Benutzung erwähnt werden. Beispielsweise könnte ein Warnhinweis zur Benutzung von rotierenden Handschuhen in der Nähe von drehenden Maschinenteilen gegeben werden. If appropriate, warnings against problems likely to be encountered or limitation of use shall be mentioned. As an example, a warning could be given about the use of tear resistant gloves used in close proximity of rotating machinery.	gegeben in englisch given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.8	Wenn die Materialien, aus denen der Handschuh besteht, ihre Leistungseigenschaften während der empfohlenen Lagerung bekanntermaßen verlieren, müssen Informationen dazu angegeben werden, um sicherzustellen, dass durch die Lagerung die Eigenschaften des Handschuhs nicht wesentlich verändert werden. If the materials constituting the gloves are known to lose their performances during recommended storage, information shall be given to ensure that the storage will not change the glove characteristics significantly.	gegeben in englisch given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.9	Wenn die geplante Leistungsfähigkeit des Handschuhs durch die Alterung bekanntermaßen erheblich beeinträchtigt werden kann, müssen die erforderlichen Angaben zur Festlegung eines angemessenen Ablaufdatums, wie nach 7.2.1.1 f) gefordert, angegeben werden. If it is known that the design performance of the glove may be significantly affected by ageing, the necessary information to establish a reasonable obsolescence date as requested in 7.2.1.1 f) shall be given.	gegeben in englisch given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.10	Bei Naturkautschuk enthaltenden Handschuhen ein Warnhinweis wie etwa „Der Handschuh enthält Naturkautschuk, der allergische Reaktionen hervorrufen kann.“ A warning for gloves containing any natural rubber, such as: "the glove contains natural rubber which may cause allergic reactions".	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.3.11	Anweisungen zum Anziehen, Ausziehen und Richten der Handschuhe, Erhalten des Komforts und der Handhygiene, Schutz vor Kontamination der Hand und gegebenenfalls Angaben zur Kombination mit anderen PSA-Elementen.	gegeben in englisch	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>



TÜVRheinland® Prüferbericht - Produkte Test Report - Products			
Prüferbericht-Nr.: Test report no.:		Seite 27 von 29 Page 27 of 29	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.12	Alle wichtigen Anweisungen zum Prüfen der Unversehrtheit des Handschuhs vor der Benutzung (z. B. Prüfen, dass der Handschuh keine Löcher, Risse, Farbveränderungen usw. aufweist und Entsorgen von Handschuhen, die solche Defekte aufweisen). Any relevant instruction to check the integrity of the glove before using it (for example check that the glove does not present holes, cracks, tears, colour change.... and discard any glove presenting such defects).	gegeben given	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.13	Anweisungen für die Lagerung. Storage instructions.	gegeben in englisch given by	P <input checked="" type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/> NT <input type="checkbox"/>
7.3.14	Wenn die Reinigung nach 4.3 angegeben ist, müssen Pflegesymbole nach ISO 3758 oder Erläuterungen sowie eine annehmbare Anzahl an Reinigungsvorgängen angegeben werden. Wenn keine Reinigung empfohlen wird, muss angegeben werden, dass der Handschuh nicht waschbar ist. Davon ausgenommen sind Einweghandschuhe. If cleaning according to 4.3 is claimed, care symbols according to ISO 3758 or explanations and an acceptable number of cleaning cycles, shall be provided. If cleaning is not recommended, it shall be indicated that the glove is not washable. This excludes single-use gloves.	n/a- Einmalhandschuh/ disposable gloves	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.3.15	Gegebenenfalls Prüfergebnisse nach 4.4 zusammen mit einer Verweisung auf die entsprechende Norm, Prüfatmosfera, Prüfmethode des Handschuhs und das angewendete Prüfverfahren bzw. die genutzte Prüfelektrode sowie die angelegte Prüfspannung nach der entsprechenden Norm. Darüber hinaus ist ein Warnhinweis anzugeben, dass die gesamte Bekleidung und alle Schuhe, die zusammen mit dieser Handschuh getragen werden, ebenfalls unter Berücksichtigung elektrostatischer Risiken getestet sein müssen. If relevant, test results according to 4.4 along with reference of corresponding standard, atmosphere for testing, area of the glove tested and test method/electrode used and the voltage applied as per the relevant standard. Moreover, a written warning shall be provided to advise that all clothing and shoes worn	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>



TÜVRheinland® Prüferbericht - Produkte Test Report - Products			
Prüferbericht-Nr.: Test report no.:		Seite 28 von 28 Page 28 of 28	
Abatz Clause	Anforderungen - Prüfungen / Requirements - Tests EN ISO 374-1:2016 + A1:2018	Messergebnisse - Bemerkungen / Measuring results - Remarks	Ergebnis Result
7.3.16	with this type of glove shall also be designed taking the electrostatic risk into account.	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.3.17	Reference to accessories and spare parts. If relevant, for example connection systems between sleeve and glove. Sofern relevant, die Art der für den Transport geeigneten Verpackung. Type of packaging suitable for transport, if relevant.	n/a	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>
7.4	Auf Nachfrage bereitzustellende Information Information to be supplied on request Eine Liste der in dem Handschuh enthaltenen Stoffe, die bekanntermaßen Allergien verursachen, siehe Anhang G, muss auf Nachfrage bereitgestellt werden, mit Ausnahme von Naturkautschuk (7.3.10). A list of the substances contained in the glove which are known to cause allergies, see Annex G, shall be supplied on request, other than natural rubber (7.3.10).	gegeben/given given	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input checked="" type="checkbox"/> NT <input type="checkbox"/>


--- Ende des Prüferberichts / End of Test Report ---



REGULATION COMPLIANCE

CE 0197

CERTIFICATE
EU Type-Examination Certificate
Regulation 2016/425/EU
Personal Protective Equipment



Registration No.:
Report No.:

Holder:


Product: Protective gloves against chemicals and micro-organisms
according to EN ISO 374-1+A1:2018 and EN ISO 374-5:2016

Identification: disposable gloves
Colour: blue
Material: nitrile, wall thickness 0,05 mm
Sizes: S (6,5) - XL (9)
Performance parameter: Type C, class 6, K: NaOH 40%
- PPE Category III - obligatory monitoring module C2 -
The EU type-examination certificate refers to the above mentioned product. This is to certify that the product complies with the essential requirements of Annex II of the regulation 2016/425/EU. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The holder is entitled to use this certificate in connection with the declaration of conformity in accordance with Annex IX.

Valid till: 07.03.2024

Date: 08.03.2019

Notified Body



Dipl.-Ing. X. Albrecht

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
Notified by Zentralstelle der Länder für Sicherheitstechnik (ZLS).

Notified under No. 0197 to the EC Commission.

CE The CE marking may be used if all relevant and effective EC Directives are complied with. CE



Business Stream Products
Textiles - PPE

TÜV Rheinland LGA Products GmbH | D-61105 Cologne Am Grauen Stein

Your correspondence:
Cornelia Albrecht
Tel. +49 221 806 5366
Mail: tip-am-st@tuev-certification@tuev.com
Cologne, 20. Juni 2022

Product surveillance acc. to Module C2 of PPE regulation (EU) 2016/425, annex VII

Your EU Type-Examination Certificate:
Product: Protective gloves against dangerous chemicals and micro-organisms
Article: Disposable gloves
Type C
Certificate of conformity
Test requirements: EN ISO 374-1:2016+A1:2018 and EN ISO 374-5:2016 (EU) 2016/425
Regulation:

Dear Mrs. Li,

The PPE Regulation (EU 2016/425) has fixed rules for the monitoring of manufactured PPE of category III with required in Annex VII (Module C2) "Conformity to type based on internal production control plus supervised product checks at random intervals".

The submitted sample of the product has been tested and in this configuration found to be in accordance with the above mentioned requirements.

Enclosed please find the test report no.

The certificate remain valid until 2024-03-07.

Date of sample picking: 2022-03-31.

Best regards

X.C. Albrecht

Certification
Signiert von: Cornelia Albrecht

X.H. D...

Lab manager
Signiert von: Miriam Dabmeier



Precisely Right.



INTERNATIONAL STANDARDS

EN 1186

Test Report No.: _____
Report Date: 11 April 2019

SUBJECT Chemical Test

TEST LOCATION TÜV SÜD China
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No 1999 Du Hu Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME _____
CLIENT ADDRESS _____

TEST PERIOD 01-Mar-2019-08-Mar-2019

TEST REQUEST In accordance with Council of Europe Res AP (2004) 4

CONCLUSION **PASS**
The submitted sample was found to comply with the overall migration requirement(s) as stated in European Resolution Res AP (2004) 4 on rubber to be used for food contact applications.

Prepared By: *Cynthia Cao*
(Cynthia Cao)
Report Drafter

Authorized By: *Leif*
(Leif)
Authorized Signatory

Chemical Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No. 1999 Du Hu Road, Minhang District
Shanghai
201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6346
Email: test.china@tuvsud.cn
Website: www.tuvsud.cn


Regional Head Office:
TÜV SÜD Certification and Testing
China Co., Ltd.
No. 101 Hong Tong Road Shanghai
200070, P.R. China

TÜV Page 1 of 2

Test Report No.: _____
Report Date: 11 April 2019

RECEIPT DATE / TEST DATE
01-Mar-2019/ 01-Mar-2019

THE FOLLOWING SAMPLE(S) WAS/WERE SUBMITTED
BY/ ON BEHALF OF THE CLIENTS AS
Sample Name: Powder free nitrile glove, blue
Sample Specification: Medium
Batch No./Date: _____
Manufacturer: /

SAMPLE NO.	DESCRIPTION	PHOTOGRAPH
	Blue glove	

TEST METHOD(S)
1. For material : Rubber
- Overall migration test for compliance with Europe Resolution Res AP (2004) 4 on rubber to be used for food contact applications.
- As specified in REGULATION (EU) No 10/2011 and its amendments, with reference to EN 1186: Part 2 (Test methods for overall migration into olive oil by total immersion) / EN 1186: Part 3 (Test methods for overall migration into aqueous food simulants by total immersion) / EN 1186: Part 14 (substitute test)

TEST RESULT(S)
1. Overall Migration Test - with reference to EN 1186: Part 2, Part 3 & Part 14

Simulant(s) Used	Test Condition	Result(s) [mg/kg]	Maximum Permissible Limit [mg/kg]
10% Ethanol	70 °C for 2 hours	<5.0	60
3% Acetic acid	70 °C for 2 hours	<5.0	60
95% Ethanol	60 °C for 2 hours	<5.0	60
Iso-octane	40 °C for 0.5 hour	<5.0	60

Note:
1. mg/kg denotes milligram per kilogram foodstuff
2. Specification is quoted from European Resolution Res AP (2004) 4 on rubber to be used for food contact applications
3. < denotes less than

Note: This report is for internal use only by the client.


-END OF THE TEST REPORT-

Chemical Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No. 1999 Du Hu Road, Minhang District
Shanghai
201108
P.R. China

Phone: +86 (21) 6037 6375
Fax: +86 (21) 6037 6346
Email: test.china@tuvsud.cn
Website: www.tuvsud.cn

Regional Head Office:
TÜV SÜD Certification and Testing
China Co., Ltd.
No. 101 Hong Tong Road Shanghai
200070, P.R. China

TÜV Page 2 of 2



REGULATION COMPLIANCE

FDA 510K

The screenshot displays the FDA's 510(k) Premarket Notification database. The page title is "510(k) Premarket Notification". The navigation menu includes Home, Food, Drugs, Medical Devices, Radiation-Emitting Products, Vaccines, Blood & Biologics, Animal & Veterinary, Cosmetics, and Tobacco Products. The search bar at the top right contains the text "Follow FDA | En Español" and a "SEARCH" button.

The main content area shows the following details for a specific 510(k) notification:

- Device Classification Name:** Polymer Patient Examination Glove
- 510(k) Number:** K171873
- Device Name:** Powder Free Nitrile Patient Examination Glove, Blue Colored, Non Sterile, Tested For Use With Chemotherapy Drugs
- Applicant:** [Redacted]
- Applicant Contact Correspondent:** [Redacted]
- Correspondent Contact:** Da Shi
- Regulation Number:** [Redacted]
- Classification Product Code:** LZA
- Subsequent Product Code:** LZC
- Date Received:** 06/23/2017
- Decision Date:** 11/15/2017
- Decision:** Substantially Equivalent (SESE)
- Regulation Medical Specialty:** General Hospital
- 510k Review Panel:** General Hospital
- Summary:** Summary
- Type:** Traditional
- Reviewed By Third Party:** No
- Combination Product:** No

At the bottom of the page, there is a note: "Page Last Updated: 06/28/2021. Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players. Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Пycкoй | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | العربية | English".



DECLARATION OF CONFORMITY

Medical

Declaration of Conformity

Manufacturer:
Address:


Product: Disposable Nitrile Examination Gloves
Designation: X-Small, Small, Medium, Large, X-Large, XX-Large

We herewith declare that the above-mentioned devices comply with the European Medical Device Regulation (EU) MDR 2017/745 and PPE Regulation (EU) 2016/425. The EU declaration of conformity is issued under the sole responsibility of the manufacturer.

By formulating the products, the chemical substances selecting was rigorous, and compliance to REACH, RoHS, Halogen-Free, SVHC 181.

We strict following the standard of U.S. and EU, no DEHP, BBP, DBP, and DIBP is using in any vinyl products.

STANDARDS
Standards Harmonized Standards applicable to this product are
EN455-1, EN455-2, EN455-3, EN455-4
EN374-1, EN374-2, EN374-3, EN374-5.

Signature: 
Date: March 02, 2021



Medical

EU Declaration of Conformity

Manufacturer:
Address:
SRN:
European Lotus NL B.V
Representative: koningin Julianaplein10, 1e Verd, 2595AA The Hague, Netherlands
SRN:
Product: Disposable Medical nitrile exam glove
X-Small, Small, Medium, Large, X-Large
GMDN Code:
UMDN Code:
Basic UDI:

Classification (MDR, Annex VIII): Class I, Rule 1.
Conformity Assessment Route: EU DECLARATION OF CONFORMITY following the Annex II + Annex III + Article 19 of MDR (EU) 2017/745.

We herewith declare that the above mentioned de products meet the transposition into national law, the provisions of the following EU Regulation and Standards. All supporting documentations are retained under the premises of the manufacturer.
is exclusively responsible for the declaration of conformity.

General applicable regulations, directives:
Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC.
Applied standards, common specification, guidance:
EN 455-1:2020, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009

Signature: _____
Date: July 06, 2021



Chemical


EU Declaration of Conformity - PPE

The manufacturer:

Representative (EU, Switzerland): CURADEN AG, Amlehnstrasse 22, 6010 Kriens, Switzerland

Declares under his sole responsibility, that the PPE reference described hereafter:

Protective nitrile Glove
PPE to be used against Category III risks
EN ISO 374-5:2016 EN ISO 374-1:2016/Type B



is in conformity with the provisions of Regulation (EU)2016/425 and with the European harmonized standards EN ISO 374-1:2016+A1:2018/Type B, EN ISO 374-5:2016, EN ISO 21420:2020 and is identical to the PPE which is subject to the EU-Type examination, under certificate number 2777, issued by the Notified Body:
SATRA Technology Europe Limited
Bracetown Business Park, Clonee
D15YN29, Republic of Ireland

and is subject to the procedure set out in Annex VII (Module C2) of the Regulation under the supervision of the Notified body:
SATRA Technology Europe Limited
Bracetown Business Park, Clonee
D15YN29, Republic of Ireland

Signature: _____
Date: January 04, 2022



Food

DECLARATION OF APPROVAL FOR THE USE OF PRODUCT WITH FOOD

The Manufacturer:

Represented by (EU, Switzerland): CURADEN AG, Amlehnstrasse 22, 6010 Kriens, Switzerland

Declaration for the following model/gloves

Nitrile VELVET GRIP blue (Premium)

Relevant category "elastomers and rubber"

meets the following regulations:

General principles of Europe Resolution Res AP(2004) 4 on rubber, EC Regulations 1935/2004 and 2023/2006 relating to good Manufacturing Practice for materials and articles intended to come into contact with food.

All ingredients, starting monomers and additives used in the production of this glove are compliant with: - all positive lists - all relevant Specific Migration Limits (SML) or restrictions of the applicable EU food legislation.



DECLARATION OF CONFORMITY

Food

Global Migration Data:

Simulants used:	Test condition	Result(s) [mg/kg]	Max. allowed limit
Ethanol (W/V) 19%	2 hours/ 70°C	<5.0	60
Acetic acid (W/V) 3%	2 hours/ 70°C	<5.0	60
Ethanol (W/V) 95%	2 hours/ 60°C	<5.0	60
Iso-octane	0.5 hours/ 40°C	<5.0	60

According to EN 1186, the analytical tolerance for simulants of foods containing water, alcohol and acids is 1 mg/dm², for simulants of fatty foods 3 mg/dm².

Storage instructions: Protect from direct sunlight, store in a cool and dry place in the original packaging. Store away from sources of ozone. If the gloves are stored properly according to the above instructions, their performance and properties will not be significantly reduced. For gloves that may be affected by aging or storage, the use-by date is stated on the packaging materials.

Signature:

Date: March 16, 2022



PACKAGING

100 GANTS
Z1 Taille L

VELVET GRIP
GANTS NITRILE jetables

• Sans latex
• Sans poudre et bleu
• Ambidextres
• Non stériles
• Gants d'examen

100 GANTS
Z1 Taille L

VELVET GRIP
NITRILE Examination GLOVES

• Disposable Gloves
• Non-sterile
• Ambidextrous
• Powder-free and blue
• Latex-free

100 GANTS
Z1 Taille L

VELVET GRIP
Einweg NITRIL-HANDSCHUHE

• Untersuchungshandschuhe
• Nicht steril
• Beidseitig
• Puderfrei und blau
• Latexfrei

100 GANTS
Z1 Taille L

VELVET GRIP
NITRILE Examination GLOVES

Importer and Distribution
CH REP CURADEN AG
Kandelstrasse 22
CH-6010 Hünen, gloves@curaden.ch

EN For medical use and treatments under hygienic conditions. Personal protective equipment. Nitrile disposable protective gloves. D Zweckbestimmung, für den ärztlichen Bereich und Behandlungen unter hygienischen Bedingungen. Persönliche Schutzausrüstung. Nitril-Einmalhandschuhe. FR Usage prévu, équipement de protection individuelle. Gant de protection jetable en nitrile. Pour le secteur médical et pour effectuer des traitements dans des conditions d'hygiène. ES Uso previsto, equipo de protección personal. Guante protector desechable de nitrilo. Para el uso médico y para realizar procedimientos en condiciones higiénicas. IT Destinazione d'uso, dispositivi di protezione individuale. Guanto protettivo monouso in nitrile. Per l'uso nel settore medico e per trattamenti medici effettuati in condizioni igieniche. RU Назначение - средства индивидуальной защиты. Назначение - одноразовые перчатки из нитрила. Для применения в сфере здравоохранения и выполнения процедур в гигиенических условиях. PL Przeznaczenie, środki ochrony osobistej. Jednorazowa rękawica nitrilowa. Do stosowania w branży medycznej i wykonywania zabiegów w warunkach higienicznych. HR Namjena, osobna zaštitna oprema. Nitrilna zaštitna rukavica za jednokratnu upotrebu. Prikladna za medicinsku podršku i likvidaciju infekcija u higijenskim uvjetima. WU Препорука, лична заштитна опрема. Нитрилна заштитна рукавица за једнократну употребу. Прикладна за медицинску подршку и ликвидацију инфекција у хигијенским условима. CZ Použití, osobní ochranné prostředky, jednorázová ochranná rukavice z nitrilu. Para utilización en el ámbito médico y en realización de procedimientos sob condiciones higiénicas. NI Bevoegd gebruik, persoonlijke beschermingsmiddelen. Nitril wegwerphandschoenen. Voor gebruik in de medische sector en het uitvoeren van behandelingen onder hygienische omstandigheden. P Utilizacja, sprzęt ochronny. Rękawiczki jednorazowe z nitrilu. Na zastosowanie w warunkach higienicznych i wykonywanie zabiegów w warunkach higienicznych. SK Účel použitia, osobný ochranný prostriedok. Jednorazová rukavica z nitrílu. Na zdravotnícku úroveň a vykonávanie ošetrovateľských a hygienických podmienok. SI Namenska uporaba, osobna zaščitna oprema. Nitrilna zaščitna rukavica za enkratno uporabo, za postopke medicinske in izvajanje tretmaja v higienskih pogojih. SE Användningsområde, personlig skyddsutrustning. Nitril engångshandskådar som är avsedda för arbete under hygieniska förhållanden. NO Tiltenkt formål, Personlig verneutrustning. Engangsplasker nitril. For bruk i helsebransjen og ved utførelse av behandlinger under hygieniske forhold. TR Kullanım amacı, kişisel koruyucu donanım. Tek kullanımlık nitril eldiven. Top katmanlı ve hijyeniği koruyan sporun tekniği ile donatılmış kulturninik cerrah teatresidir. RO Destinație, echipament de protecție individuală. Mănușă de protecție de unică folosință din nitril. Pentru utilizare în domeniul medical și efectuarea tratamentelor în condiții corect organizate de igienă.

4 260740 980185

EC REP LOTUS NL B.V., Koningsh Julianaplein 10,
1e Verd. 2535AA The Hague, Netherlands
Tel: +31-6-41-63999

For user information/certificate please visit:
www.curaden.com/en/ve-gloves
For more information see website

ISO 13485:2016 ISO 9001:2015 EN ISO 21420:2020 MDR
EN ISO 374-1:2016+A1:2018/Type B Level EN ISO 374-4:2019
Degradable %

40% Sodium hydroxide (K)	6	4.3%
32% Hydrogen peroxide (P)	2	24.1%
27% Formaldhyde (T)	4	34.3%

MADE IN PRC

EN14931
MD
AQL 1.5
ENT16
EN853 1,2,3,4

CUR_VG_Z1_CERT_LN_EN_30012023_v32

22 of 22